File: acorde_status.doc
Last updated: 28 February 2008

Readout sub-detector status

ACORDE

Local Trigger Crate: **C26B**

Table of contents

L0
BUSY
TTC optical connections
Prepulse
L1 Message format

Trigger Questionnaire status: Fully up-to-date
Questionnaire last updated: 9 February 2008

L0

L0 over TTC option (Yes/No): **Yes**

BUSY

BUSY cable inputs

Number of BUSY connections: **1**

Input BUSY1:
- Enable (Yes/No): **Yes**
- Cable number: **18789**
- Cable length: **40m**

Input BUSY2:
- Enable (Yes/No): **No**

Fanin board (Yes/No): **No**

TTC optical connections

Number of optical connections: **1**
Length of the optical patchcord: **2.2m**
Length of the optical cable: : **40m**
Optical coupler used (Yes/No): **No**
Optical patch-panel slot: **1**
Optical patch-panel connector: **1**

Prepulse

Prepulse used (Yes/No): **No**
Delay setting D (TTCvi): **3422**
Lead time to the calibration L0 [(3556-D-54) x 25ns]: **2µs**

L1 Message format

(First word only/All five words): **All five words**