File: mtrg_status.doc
Last updated: 20 September 2007

Readout sub-detector status

MTRG

Local Trigger Crate: **C24T**

Table of contents

L0
BUSY
TTC optical connections
Prepulse
L1 Message format

Trigger Questionnaire status: Fully-up-to-data
Questionnaire last updated: 30 January 2007

L0

L0 over TTC option (Yes/No): **No**

L0 cable outputs

Number of L0 connections: **2**

Output $L0_1$:
- Cable number: **3148**
- Cable length: **10m**

Output $L0_2$:
- Cable number: **3149**
- Cable length: **10m**

Fanout board (Yes/No): **No**

BUSY

BUSY cable inputs

Number of BUSY connections: **2**

Input BUSY1:
- Enable (Yes/No): **Yes**
- Cable number: **3150**
- Cable length: **10m**

Input BUSY2:
- Enable (Yes/No): **Yes**
- Cable number: **3151**
- Cable length **10m**

Fanin board (Yes/No): **No**

TTC optical connections

Number of optical connections: **2**
Length of the optical patchcord: **1.8m**
Length of the optical cable: **10m**
Optical coupler used (Yes/No): **No**
Optical patch-panel slot: **5**
Optical patch-panel connector(1, 2, 3): **1, 2**

Prepulse

Prepulse used (Yes/No): **Yes**
Delay setting D (TTCvi): **3422**
Lead time to the calibration L0 [(3556-D-54) x 25ns]: **2µs**

L1 Message format

(First word only/All five words): **First word only**